

Issued by NMI Certin UK (TIC) Ltd.,
appointed and approved by the Secretary of State to perform tasks with respect to conformity modules mentioned in regulation 39 of the Measuring Instruments Regulations 2016 as amended, after having established that the measuring instrument meets the applicable requirements of the Measuring Instruments Regulations 2016 as amended, to:

Manufacturer **MeterSit**
Viale dell'Industria, 31
35129 Padova
Italy

Measuring instrument **Thermal-mass flow gas meter**
Type : x4850xx
x4853xx
x4854xx

Manufacturer's mark or name : MeterSit

Destined for the measurement of : Gas volume of natural gas or hydrogen,
See description §1.2.2

Accuracy class : Class 1,5

Environment classes : M1 / E2

Temperature range : -25 °C / +55 °C

Designed for : Condensing humidity

Location : Open

Further properties are described in the annexes:

- Description T10362-UK revision 4;
- Documentation folder T10362-UK-4.

Valid until 26 September 2032

Remark: This revision replaces the previous versions, including its documentation folder.

Issuing Authority **NMI Certin UK (TIC) Ltd., Approved Body number 8506**
10 February 2023

Certification Board

NMI Certin UK (TIC) Ltd.
5 Cecil Pashley Way
Shoreham Airport
Shoreham-by-Sea
West Sussex, BN43 5FF
United Kingdom
T +44(0)7432156880
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The appointment of NMI Certin UK (TIC) Ltd., as Approved Body can be verified at <https://www.gov.uk/uk-market-conformity-assessment-bodies>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



23275